Summary

WORD COUNT	CHARACTER COUNT
3317 Words	18594 Characters
PAGE COUNT	FILE SIZE
6 Pages	302.5KB
SUBMISSION DATE	REPORT DATE
Oct 23, 2022 8:20 AM GMT+7	Oct 23, 2022 8:20 AM GMT+7

14% Overall Similarity

The combined total of all matches, including overlapping sources, for each database.

- 5% Internet database
- Crossref database
- 9% Submitted Works database

Excluded from Similarity Report

- Bibliographic material
- Manually excluded sources

- 10% Publications database
- Crossref Posted Content database
- Cited material

AUTHOR

3758_Artikel_Polimesin_S2_Fauzi_Irza_A Irza Sukmana

PAPER NAME

ug2021.pdf



JURNAL POLIMESIN Unit Penelitian dan Pengabdian Kepada Masyarakat (P2M) Politeknik Negeri Lhokseumawe

ISSN Print: 1693-5462, ISSN Online: 2549-1199 Website: http://e-jurnal.pnl.ac.id/index.php/polimesin

Application of magnesium alloys in orthopedic implant

Fauzi Ibrahim, Irza Sukmana^{*}

Department of Mechanical Engineering, Faculty of Engineering, Universitas Lampung, Bandar Lampung, 35143, Indonesia

* Corresponding author: irza.sukmana@eng.unila.ac.id

Abstract

Technological updates in the field of advanced materials research are now and to focuses on biomedical materials application and utilization of Magnesium and its' alloys. Magnesium (Mg) has been widely studied as an alternative material for biodegradable orthopedic implant applications. Recent studies regarding the potential application of Mg have been done related to its' mechanical properties, biodegradation profile, and the *a-vitro* and *in-vivo* testing. This study aims to review the Mg properties, production process, biomaterial roadmap, and the concern of chemical composition of Mg alloy in orthopedic application. Future potential improvement of the magnesium alloys properties is also highlighted.

Keywords: Magnesium alloys; biodegradable; orthopedic implant; biomaterial roadmap

1. Introduction.

Orthopedic implants are now closely related to the concept of developing biomaterials, especially in human bones. Human bones are very susceptible to pain, dislocation and fracture. In general, fractures that occur can be caused by external factors such as accidents. However, it is possible for fractures to be caused by disease or conditions of bone abnormalities from birth. Bone fractures are generally caused by excessive external loads or forces applied to the bone [1]. Orthopedic biomaterials can be implanted into the body as an alternative to healing or replacing damaged or lost tissue. The implantation process is generally done with two surgeries when the second surgery found problems such as infection due to the degradation of the implant.

The main requirements for implant material candidates must be degraded, adapt to their biological environment and have adequate mechanical properties. Mechanical properties for implant materials in terms of yield strength, Young's modulus, wear resistance and tensile or compressive strength. Another characteristic of implant materials is that they have a relatively light density and low manufacturing costs. Observation of implant materials that have biodegradable properties is needed to improve performance, function and to achieve the goal of tissue regention [2].

There are three main types of materials for implants: ceramic-based, metals and polymers. Stainless steel and titanium alloys are kind of metallic materials that have been widely used in orthopedic implants [3]. The biggest drawback of those implant materials is that they are not completely degraded, and there must be a second surgery after the bone is declared healed. Most researchers focus on the development of implant material that have biodegradable properties. Therefore, magnesium and its alloys are the best candidates for implant materials. Magnesium generally contains alloying elements: aluminum, zinc and manganese [4].

density, a Young's modulus that is close to bone morphology with high initial stability. This review aims to provide an overview of research on magnesium alloys as orthopedic implants with more descriptive mechanical characteristics, factors and costs.

The effect of alloying elements or chemical composition on biodegradation effects and mechanical performance will be discussed in detail. The expectation of this review is to help research of implant materials based on magnesium alloys.

2. Magnesium for orthopedic implant.

The development and use of magnesium as an implant have been around for more than a century. Formerly, the materials applied for biomaterials in particular were materials found in industry or society and not special materials, after that the material was modified and tested for biocompatibility.

There are three types of biomaterial implants such as ceramics, metals and polymers. Metallic materials have been widely used in orthopedic applications such as Co-Cr, titanium, and stainlesssteel alloys. Disadvantage of metallic biomaterials is that they are not properly degraded in the body and require secondary surgery after the bone heals; as presented on Table 1.

Table 1. Advantages, disadvantages and applications of implant materials							
No.	Material	Advantages	Disadvantages	Applications	References		
1	316L Stainless Steel	 Easily Available Low cost Accepted biocompatibilit y and toughness Easily fabricate 	 High modulus Poor corrosion resistant Poor wear - resistant Elements can cause allergies 	¹ Jone plates, bone screws and pins, wires.	[5,6]		
2	Co-Cr 8 Illoys	 Resistant to corrosion, fatigue and wear High strength Long term biocompatibility 	 Expensive Difficult to fabricate High modulus Toxic when Co, Cr and Ni ions are released 	Shorter term implants-Bone plates and wires, Total hip replacements PHR)-Stem or hard-on-hard bearing system	[7]		
3	Mg Alloys (solid)	 Biocompatible Biodegradable Bioresorbable Density and Young's modulus similar to bone Light Similar to tibial and cortical bone 	 Low corrosion resistance Control of hydrogen gas must be carried out during the degradation process 	Bone screws, bone plates, bone pins	[8-10]		
4	Mg Alloys (porous)	 Low cost Easily to control pore formation The product is close to the mold in shape Similar to cancellous bone for bone scaffold applications 	 Low corrosion resistance Ratio of the composition must be balanced Pores resulted are sometimes different 	Bone screws, bone plates, bone pins	[11,12,13]		

³Magnesium and its alloys also have a good biocompatibility, ²⁵magnesium is a very abundant element in the human body. Since the beginning of 1878, magnesium and its alloys have been studied for the utility of modern or advanced materials as biomaterials [14,15].

However, developments only started in the beginning of 2000s, such as corrosion control, and controlled degradation time to provide a healing process without surgical removal of the implant. The toxicity percentage of magnesium alloying elements is one of the main disadvantages of magnesium alloys.

The selection process of alloying elements must be very careful to optimize the revolutionary magnesium alloys in the future. The density and Young's modulus of magnesium are in the same range as natural human bone, making it an alternative for orthopedic implant applications.

The Young's modulus of magnesium is 40 - 50 GPa [16]. This value tends to resemble the stiffness of human bone, which is around ~40 GPa.

Theoretically, Magnesium has a tensile strength unlike the steel, but magnesium has the properties and abilities when used as a human bone implant material, not only safe but reducing costs significantly compared to the price of other implant materials used in the market now.

The selection for biomedical application of magnesium is not only based on its mechanical properties, but also related to its' degradation and biocompatibility that mostly depend on the alloying elements.

Additional alloying elements in Mg may improve the mechanical properties through the development of hexagonal closed-pack (HCP) matrix and precipitation hardening [17,18].

2.1. Production Process

Powder metallurgy is a synthesis technique in the solid state, followed by a sintering process with an adjustable temperature but tends to be expensive with some risks such as the difficulty of transporting metal in powder form and needing a mold. Casting is the most economical way of production magnesium but there are high risks such as porosity, poor mechanical properties and high impact on the environment [19].

The process of production degradable magnesium is classified into 2 conditions, they are solid or liquid conditions. In the molten state, magnesium can be formed by casting; stir casting, squeeze casting and disintegrated melt deposition, while the solid condition was formed by a powder metallurgy process [20,21]. The type of Mg alloys, production process, chemical 16 mposition, and their applications in biomedical is snown in table 2.

Table 2. Types, formation, chemical composition, and applications of magnesium alloys

No	Alloy	Production	_	Chemical Composition				Applications	
INO.	Туре	Process	Al	Zn	Mn	Si	RE	Zr	Applications
1	AM60	Die Casting	6		>0,13				Bioresorbable Stent
2	AZ31B	Extrusion and rolling	3	1	0,3				Bone screws, bone plates, bone pins
3	AZ80A	Extrusion	8	0,5	0,2				Bone screws, and pins
4	HK31A	Rolling						0,7	Bone plates
5	AZ91B	Die Casting	9	0,7	>0,13				
6	AZ91D	Die Casting	9	0,7	0,2				Long term
7	EZ33A	Sand Casting		3			3	0,8	biocompatibility
8	AS41A	Die Casting	4		0,3	1	0,3		

The hot or cold working metal is divided into two parts: rolling and extrusion. Rolling is a roduction process that uses a pair of rollers in order to reduce the thickness of the material by compressing it to a certain size with hot or cold conditions, while extrusion of metal production uses the compressive force of a mold and produces metal that has been reduced or has reduced crosssection

The production process is expected to be easy, along with the development of current biomaterial technology. The potential application of magnesium alloys for orthopedic implant applications has not been said perfect, because there are several problems such as controlling aydrogen gas during the degradation process, the level of toxicity that must be identified, corrosion and mechanical behavior that must be anticipated before in-vitro, also studied during in-vivo.

2.2. Alloy Elements and Mechanical Properties

The elements of magnesium that are commonly used are including: aluminum (Al), zinc (Zn), manganese (Mn), and calcium (Ca). Elements can be classified into three types, they are toxic elements, nutritional elements and elements that cause allergies. Cadmium, beryllium, lead, barium, thorium, and iron are some examples for clumps of toxic elements. The elements contained in the alloy have their respective roles in human bones. Magnesium is a mineral that is needed by the human body and is evenly distributed in tissues, blood, body fluids, and other organs [22].

Magnesium has been shown to promote the process of bone establishment after being damaged. Previous research has shown that calcium plays an important role in the process of enhancing bone mineral density, strength, stretch ability, and calcium also gives bone hard properties for forming bone structure to support the human body. The zinc element in the alloy acts as a reinforcement and to overcome corrosive. The density and Young's modulus of zinc is slightly higher than that of magnesium, this will delay the healing process. The role of manganese as a stimulant of bone metabolism and it is able to create bone-forming enzymes, and bo can prevent osteoporosis [23].

Alto other alloying elements that may contribute to improve the mechanical properties of Mg are including: aluminum, zinc, nickel and copper can increase the value of strength and ductility. Elements such as lead on increase its strength but decrease its ductility. The addition of alloying elements such as nickel and iron should be as minimal as possible.

Nickel and iron are considered as unnecessary substances because they cause toxicity to the implant material. The alloy not only supports its mechanical properties, but also must be degraded over time during the implantation process. Research that focuses on corrosion rate and density has been done to obtain efficient results. Alloys of magnesium and calcium have mechanical properties and biocompatibility similar to cortical bone, the addition of calcium to magnesium can suppress the rate of oxidation during the casting process and improve the mechanical properties, also corrosion resistance according to the percentage of calcium.

Zinc and nickel can reduce toxic et ments such as iron, thereby increasing the value of tensile strength, hardness, ductility, and corresion resistance. However, high levels of zinc will form pores and a eutectic phase which reduces the strength value. Total percentage of lead is less than 5% will mprove the strength and ductility of magnesium.

alloys the mechanical properties of magnesium epend on the percentage of alloying elements, grain size, and intermetallic phase formed. The orthop dic implant material must have sufficient strength to support the bone structure during the healing process. The yield strength and maximum strength of magnesium are obtained from tensile testing. The series of stages of magnesium biocompatibility testing must be done in stages according to the biomaterial roadmap.

3. Roadmap of Mg Alloy

Biomaterials considers as materials that produce or synthesized for the purpose of applications related to contact with the human body. Biomedical material is aims to repair, replace, support, and to restore the function of organs or the human body. It can be in the form of materials with special technical specifications or in the form of some medical devices and biomaterials also have other purposes such as producing a product that is biocompatible, economical and produced easily in order to reduce very expensive costs.

Biomaterials are also one of the hopes for alternative biomedical materials that have high potential to be developed, particularly orthopedic implant materials [24]. The material which will be used as implant material will be modified and tested for biocompatibility with different sub-fields of science

In the field of materials science and engineering, the implant material is tested for the first stage, in the in vitro Lab. At this stage, testing of strength, hardness, microstructure, fatigue, density, porosity, and corrosion rate of the candidate materials have to be done. Also, the material must be characterized related to the targeted implant site. For example, Mg may be tested as solid materials for the bone implant purposes, as well as in porous structure for bone scaffolding material. Following the success of the in vitro test, the second stage is the animal or in vivo testing.

One this second stages, the biocompatibility of implant material should be tested using a scarifying animal. This in vivo test needs to be handled by animal husbandry doctors and will be carried out in the group of knowledge of tissue engineering and animal cell culture technology.

Once its' results show a biocompatibility and have no rejection by the animal the further test may be runs using a clinical test. to determine the biocompatibility level of the material inside the human's body.

This clinical testing should be runs by the medical doctors in the respected research hospital. The testing of biomaterial candidate inside the human body is the last step before the approval may be given to the biomaterial candidate [25].

Conclusions Magnesium and its alloys have the potential to be used as orthopedic implant materials. The alloying elements which are used improving the mechanical properties of magnesium need attention because the low or high percentage of alloying elements greatly affects the biocompatibility of magnesium. Cost factors, risks, and the process of production magnesium which will be used as an alternative implant material must follow the standards that have been tested or have been carried out in previous research. This review is expected to highlight the potential of Magnesium and its' alloying elements as orthogoic implant material. The production processes, mechanical properties and the biocompatibility of Mg and its' alloys open a potential application as biomedical material for orthopedic implants.

Acknowledgement

This study is supported by Hibah Penelitian Kerja sama Internasional DIPA FT Universitas Lampung, International Cooperative Research grant of FT UNILA No. 4062/UN26.15/LK.03/2021.

References

Li, Y., Wen, C., Mushahary, D., Sravanthi, [1] R., Harishankar, N., Pande, G. and Hodgson, P., Mg-Zr-Sr alloys as biodegradable implant materials. Acta biomaterialia, 8(8), pp.3177-3188. 2012.

- [2] Chou, D.T., Hong, D., Saha, P., Ferrero, J., Lee, B., Tan, Z., Dong, Z. and Kumta, P.N., In vitro and in vivo corrosion, cytocompatibility and mechanical properties of biodegradable Mg–Y–Ca–Zr alloys as implant materials. *Acta biomaterialia*, 9(10), pp.8518-8533. 2013.
- [3] Wen, C., Guan, S., Peng, L., Ren, C., Wang, X. and Hu, Z., Characterization and degradation behavior of AZ31 alloy surface modified by bone-like hydroxyapatite for implant applications. *Applied Surface Science*, 255(13-14), pp.6433-6438. 2009.
- [4] Hong, D., Saha, P., Chou, D.T., Lee, B., Collins, B.E., Tan, Z., Dong, Z. and Kumta, P.N., In vitro degradation and cytotoxicity response of Mg–4% Zn–0.5% Zr (ZK40) alloy as a potential biodegradable material. *Acta biomaterialia*, 9(10), pp.8534-8547. 2013.
- Oshkour, A. A., Pramanik, S., Mehrali, M., [5] Yau, Y. H., Tarlochan, F., & Osman, N. A. A. Mechanical and physical behavior of developed functionally newly graded materials and composites of stainless steel 316L with calcium silicate and hydroxyapatite. Journal of the mechanical behavior of biomedical materials, 49, pp. 321-331. 2015.
- [6] Karamian, E., Motamedi, M. R. K., Khandan, A., Soltani, P., & Maghsoudi, S. An in vitro evaluation of novel NHA/zircon plasma coating on 316L SS dental implant. *Progress* in Natural Science: Materials International, 24(2), pp. 150-156. 2014.
- [7] Delaunay, C., Petit, I., Learmonth, I. D., Oger, P., & Vendittoli, P. A. Metal-on-metal bearings total hip arthroplasty: the cobalt and chromium ions release concern. *Orthopaedics & Traumatology: Surgery & Research*, 96(8), pp. 894-904. 2010.
- [8] Zeng, R.C., Cui, L.Y., Jiang, K., Liu, R., Zhao, B.D. and Zheng, Y.F., In vitro corrosion and cytocompatibility of a microarc oxidation coating and poly (l-lactic acid) composite coating on Mg–1Li–1Ca alloy for orthopedic implants. ACS Applied Materials & Interfaces, 8(15), pp.10014-10028. 2016.
- [9] Levorova, J., Dugova, L., Ulmann, D., Vrbova, R., Duskova, J., & Foltan, R. In vivo biodegradation of magnesium alloys screws in rabbit tibia: influence on bone healing. *International Journal of Oral and Maxillofacial Surgery*, 46, pp. 347. 2017.
- [10] Chaya, A., Yoshizawa, S., Verdelis, K., Myers, N., Costello, B. J., Chou, D & Sfeir, C. In vivo study of magnesium plate and screw degradation and bone fracture healing. *Acta biomaterialia*, 18, pp. 262-269. 2015.
- [11] Yusop, A.H.M., Alsakkaf, A., Kadir, M.R.A., Sukmana, I. and Nur, H., Corrosion of porous Mg and Fe scaffolds: a review of mechanical and biocompatibility responses. *Corrosion*

Engineering, Science and Technology, pp.1-17. 2021.

- [12] Noor, J., Djuansjah, J.R.P., Kadir, M.R.A., and Sukmana, I., Porous Magnesium Scaffolds for Bone Implant Applications: A Review. Advanced Materials Research. 2015.
- [13] Zhuang, H., Han, Y. and Feng, A., Preparation, mechanical properties and in vitro biodegradation of porous magnesium scaffolds. *Materials Science and Engineering: C*, 28(8), pp.1462-1466. 2008.
- [14] Prakash, C., Singh, S., & Ramakrishna, S. Characterization of indigenously coated biodegradable magnesium alloy primed through novel additive manufacturing assisted investment casting. *Materials Letters*, 275, 128-137. 2020.
- [15] Hart, N. H., Nimphius, S., Rantalainen, T., Ireland, A., Siafarikas, A., & Newton, R. U. Mechanical basis of bone strength: influence of bone material, bone structure and muscle action. *Journal of musculos-keletal & neuronal interactions*, 17(3), 114. 2017.
- [16] Rodrigues, A., Caetano-Lopes, J., Nery, A., Sousa, E., Polido-Pereira, J., Vale, M., & Canhão, H. Evaluation of Bone Mechanical Strenght and Fracture Risk Assessment [Frax] In Patients with Hip Joint Replacement Surgery. Acta reumatologica portuguesa, 34(3). 2009.
- [17] Faruk, M. E. R. T. Wear behaviour of hot rolled AZ31B magnesium alloy as candidate for biodegradable implant material. *Transactions of Nonferrous Metals Society of China*, 27(12), pp. 2598-2606.2017.
- [18] Zhang, L. C., Xu, M., Hu, Y. D., Gao, F., Gong, T., Liu. Biofunctionization of biodegradable magnesium alloy to improve the in vitro corrosion resistance and biocompatibility. *Applied Surface Science*, 451, pp. 20-31. 2018.
- [19] Friedrich, Horst E., Barry L. Mordike. 2006. Magnesium Technology, Metallurgy, Design Data, Applications. Germany. Springer.
- [20] Buldum, B. B., Aydın, S. I. K., & Ozkul, I. Investigation of magnesium alloys machinability. *International Journal of Electronics Mechanical and Mechatronics Engineering*, 2(3), pp. 261-268. 2013.
- [21] Witte, F., Hort, N., Vogt, C., Cohen, S., Kainer, K. U., Willumeit, R., & Feyerabend, F. Degradable biomaterials based on magnesium corrosion. *Current opinion in solid state and materials science*, 12(5-6), pp. 63-72. 2008.
- [22] Witte, F., Kaese, V., Haferkamp, H., Switzer, E., Meyer-Lindenberg, A., Wirth, C. J., & Windhagen, H. In vivo corrosion of four magnesium alloys and the associated bone response. *Biomaterials*, 26(17), pp. 3557-3563. 2005.

- [23] Jayasathyakawin, S., Ravichandran, M., Baskar, N., Chairman, C. A., & Balasundaram, R. Mechanical properties and applications of Magnesium alloy–Review. *Materials Today: Proceedings*, 27, pp. 909-913. 2020.
- [24] I. Sukmana. *Ilmu dan Teknologi Biomaterial*. Yogyakarta Indonesia: Teknosain, 2017. pp. 48-57.
- [25] I. Sukmana, S. Savetlana, Y. Burhanuddin, M.A. Wicaksono, and H. Nur. "Fabricating and Testing of Porous Magnesium Through Powder Metallurgy Technique using TWSH for Biodegradable Bone Scaffold Material". *Journal of Engineering and Scientific Research*, vol. 1. No. 2, pp. 78-83, Dec. 2019.

• 14% Overall Similarity

Top sources found in the following databases:

- 5% Internet database
- Crossref database
- 9% Submitted Works database
- 10% Publications database
- Crossref Posted Content database

TOP SOURCES

The sources with the highest number of matches within the submission. Overlapping sources will not be displayed.

1	R Radha, D Sreekanth. "Insight of magnesium alloys and composites for Crossref) 2%
2	De Montfort University on 2021-09-07 Submitted works	1%
3	jurnal.upnyk.ac.id	<1%
4	coek.info Internet	<1%
5	"Magnesium Technology 2017", Springer Nature, 2017 Crossref	<1%
6	scindeks-clanci.ceon.rs Internet	<1%
7	American University in Cairo on 2018-12-18 Submitted works	<1%
8	Bahrain Training Institute on 2018-03-17 Submitted works	<1%

Mohammad Asgari, Ruiqiang Hang, Chang Wang, Zhentao Yu, Zhiyon Crossref	^{ig} <1%
Cijun Shuai, Yuanzhuo Zhou, Youwen Yang, Chengde Gao, Shuping Po Crossref	en <1%
Jingyao Wu, Daoli Zhao, John Ohodnicki, Boeun Lee et al. "In vitro an Crossref	d i <1%
Myrissa, Anastasia, Nezha Ahmad Agha, Yiyi Lu, Elisabeth Martinelli, Crossref	J <1%
University of Nottingham on 2016-12-13 Submitted works	<1%
University of Strathclyde on 2012-04-25 Submitted works	<1%
zone.biblio.laurentian.ca	<1%
"Biomaterials in Orthopaedics and Bone Regeneration", Springer Scie	<1%
A. Torkian, G. Faraji, M. S. Pedram. "Mechanical properties and in viv Crossref	^o <1%
Maulana Azad National Institute of Technology Bhopal on 2019-08-2 Submitted works	1 <1%
The University of Manchester on 2011-12-09 Submitted works	<1%
University of Birmingham on 2013-03-08 Submitted works	<1%

21	University of Birmingham on 2020-02-18 Submitted works	<1%
22	docplayer.net Internet	<1%
23	encyclopedia.pub Internet	<1%
24	gyan.iitg.ac.in Internet	<1%
25	revistas.unal.edu.co Internet	<1%
26	mdpi.com Internet	<1%
27	Irza Sukmana, Fauzi Ibrahim, Mohammad Badaruddin, Hadi Nur. "Low	<1%